



Certificate of Analysis

RAPICIDE® PA

Part B

Product Code: ML02-0115
Lot Number: 012116
Expiration Date: 2017/11

Components	Test Method	Specification	Analysis
Sodium	20100-070	743.0 - 909.0 mEq/L	775.5 mEq/L
Benzotriazole	20601-489	0.9 – 1.1% w/w	1.0 %w/w
pH	20002-154	11.4 to 12.4 as is	11.7

This lot has been reviewed and meets specified requirements.

Quality Analyst

31May2016

Analysis Date

Quality Assurance

31May2016

Release Date

Medivators Inc.

14605 28th Avenue North
Plymouth, Minnesota 55447
Telephone: (763) 553-3300
Fax (763) 553-3387

RAPICIDE® is a registered trademark of Medivators Inc.