



Certificate of Analysis

RAPICIDE® PA

RTU Part B

Lot Number: 012217
Expiration Date: 2018/10
Quality Control Report: 170407

Components	Specifications	Analysis
Na+	35.5 – 43.4 mEq/L	42.0
Benzotriazole	0.05% ± 0.005%	0.046
pH, (as is)	10.6 – 11.6 (as is)	11.6

This lot has been reviewed and meets specified requirements.

Quality Analyst

21 April 2017

Analysis Date

Quality Assurance Pa

24 April 2017

Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com