



## Certificate of Analysis

# RAPICIDE® PA

### Part A

Product Code:

**ML02-115**

Lot Number:

012918

Expiration Date:

2019/09

| Components                    | Specifications     | Analysis |
|-------------------------------|--------------------|----------|
| H <sub>2</sub> O <sub>2</sub> | 20.0 – 24.0% (w/w) | 23.6 %   |
| PAA                           | 4.0 – 6.0 % (w/w)  | 5.6 %    |
| Acetic Acid                   | 8.0 – 10.0% (w/w)  | 8.8 %    |
| pH, (as is)                   | 0.5 – 1.1 (as is)  | 0.7      |

This lot has been reviewed and meets specified requirements.

Quality Analyst

27 March 2018

Analysis Date

Quality Assurance

Release Date

Medivators Inc.  
14605 28th Avenue North  
Minneapolis, MN 55447 USA  
Telephone: (763) 553-3300  
Fax (763) 553-3387  
[www.medivators.com](http://www.medivators.com)

RAPICIDE® is a registered trademark of Medivators Inc.