



## Certificate of Analysis

# RAPICIDE® PA

### Part A

Product Code: ML02-115  
Lot Number: 012919  
Expiration Date: 2020/09

| Components                    | Specifications     | Analysis |
|-------------------------------|--------------------|----------|
| H <sub>2</sub> O <sub>2</sub> | 20.0 – 24.0% (w/w) | 23.0 %   |
| PAA                           | 4.0 – 6.0 % (w/w)  | 5.6 %    |
| Acetic Acid                   | 8.0 – 10.0% (w/w)  | 8.8 %    |
| pH, (as is)                   | 0.5 – 1.1 (as is)  | 0.8      |


This lot has been reviewed and meets specified requirements.

FDC

Quality Analyst

28-MAR-19

Analysis Date

  
Quality Assurance

01-APR-19

Release Date

Medivators Inc.  
14605 28th Avenue North  
Minneapolis, MN 55447 USA  
Telephone: (763) 553-3300  
Fax (763) 553-3387  
[www.medivators.com](http://www.medivators.com)

RAPICIDE® is a registered trademark of Medivators Inc.