



Quality Control Report and Certificate of Analysis


Formula: SB-144

Lot #: 012A08

	<u>Specification</u>	<u>Analysis</u>	<u>Deviation</u>
Sodium	<u>4500</u> mEq/L	<u>4488.8</u> mEq/L	<u>-0.25%</u>
Calcium	<u>101.25</u> mEq/L	<u>97.6</u> mEq/L	<u>-3.56%</u>
Potassium	<u>135</u> mEq/L	<u>134.2</u> mEq/L	<u>-0.57%</u>
Magnesium	<u>33.75</u> mEq/L	<u>32.8</u> mEq/L	<u>-2.96%</u>
Chloride	<u>4770</u> mEq/L	<u>4800.3</u> mEq/L	<u>0.63%</u>
Dextrose	<u>9000</u> mg/dL	<u>9100.0</u> mg/dL	<u>1.11%</u>
Conductivity 25 C	<u>172.5</u> mS/cm	<u>167.8</u> mS/cm	<u>-2.72%</u>
Pyrogen Test	<u>Non-pyrogenic</u>	<u>Non-pyrogenic</u>	
pH Conc.	<u>1.8 - 3</u> pH Units	<u>1.96</u> pH units	
pH 1:34	<u>3 - 4</u> pH Units	<u>3.39</u> pH units	

Approved for Release:

20600-379/D


Quality Control

01-03-08
Date