

## *Certificate of Conformance*

# **McKESSON® OPA/28 Test Strips**

**Part Number:**

**73-OPA28B**

**Kit Lot Number:**

**014045**

**Vial Lot Number:**

**020045**

**Expiration Date:**

**12/16**


### **Quality Control Report:**

MEDIVATORS certifies that the test strips indicate a positive or pass reaction with a solution mixed to a concentration of 0.50%  $\pm$  0.05% OPA.

MEDIVATORS certifies that the test strips indicate a negative or fail reaction with a solution mixed to a concentration of 0.35%  $\pm$  0.05% OPA.

The statistical analysis and approval test procedure is on file and available for review.

This lot has been reviewed and meets specified requirements.

  
\_\_\_\_\_  
Quality Assurance

4/22/2015  
Release Date

**MEDIVATORS Inc.**  
14605 28<sup>th</sup> Avenue N  
Minneapolis, MN 55447  
Tel: 800.444.4729  
Fax: 763.553.3387  
[www.medivators.com](http://www.medivators.com)