

Certificate of Conformance

MaxiCide® OPA 28 Test Strips

Part Number: **112-6977**

Kit Lot Number: **014077**

Vial Lot Number: **044077**

Expiration Date: **2019-07-31**

Quality Control Report:

MEDIVATORS certifies that the test strips indicate a positive or pass reaction with a solution mixed to a concentration of 0.50% \pm 0.05% OPA.

MEDIVATORS certifies that the test strips indicate a negative or fail reaction with a solution mixed to a concentration of 0.35% \pm 0.05% OPA.

The statistical analysis and approval test procedure is on file and available for review.

This lot has been reviewed and meets specified requirements.



Linda Heidtner
Quality Assurance

Medivators Inc.
14605 28th Avenue N
Minneapolis, MN 55447
Tel: 800.444.4729
Fax: 763.553.3387
www.medivators.com

8/18/2017

Release Date