



Certificate of Conformance

RAPICIDE® OPA/28 Test Strips

Part Number:

ML02-0137

Lot Number:

015024

Vial Lot Number:

007024

Expiration Date:

08/15

Quality Control Report:

MEDIVATORS certifies that the test strips indicate a positive or pass reaction with a solution mixed to a concentration of $0.50\% \pm 0.05\%$ OPA.

MEDIVATORS certifies that the test strips indicate a negative or fail reaction with a solution mixed to a concentration of $0.35\% \pm 0.05\%$ OPA.

The statistical analysis and approval test procedure is on file and available for review.

This lot has been reviewed and meets specified requirements.

Quality Assurance

4/8/2014

Release Date

MEDIVATORS Inc.

14605 28th Avenue N

Minneapolis, MN 55447

Tel: 800.444.4729

Fax: 763.553.3387

www.medivators.com