

Certificate of Conformance

McKESSON® OPA/28 Test Strips

Part Number:

73-OPA28B

Kit Lot Number:

015055

Vial Lot Number:

022055

Expiration Date:

01/17

Quality Control Report:

MEDIVATORS certifies that the test strips indicate a positive or pass reaction with a solution mixed to a concentration of 0.50% \pm 0.05% OPA.

MEDIVATORS certifies that the test strips indicate a negative or fail reaction with a solution mixed to a concentration of 0.35% \pm 0.05% OPA.

The statistical analysis and approval test procedure is on file and available for review.

This lot has been reviewed and meets specified requirements.

Erin A. Grotto
Quality Assurance

5/28/2015
Release Date

MEDIVATORS Inc.
14605 28th Avenue N
Minneapolis, MN 55447
Tel: 800.444.4729
Fax: 763.553.3387
www.medivators.com