

Certificate of Analysis

RAPICIDE® PA

Part B

Product Code: **ML02-0115**

Lot Number: **015216**

Expiration Date: **2017/12**

Components	Test Method	Specification	Analysis
Sodium	20100-070	743.0 - 909.0 mEq/L	801.8 mEq/L
Benzotriazole	20601-489	0.9 – 1.1% w/w	1.0 %w/w
pH	20002-154	11.4 to 12.4 as is	11.5

This lot has been reviewed and meets specified requirements.


Quality Analyst

07Jul2016
Analysis Date


Quality Assurance

07Jul2016
Release Date

Medivators Inc.
14605 28th Avenue North
Plymouth, Minnesota 55447
Telephone: (763) 553-3300
Fax (763) 553-3387

RAPICIDE® is a registered trademark of Medivators Inc.