



Certificate of Analysis

RAPICIDE™ PA

RTU Part A

Lot Number: 015419
Expiration Date: 2020/10

Components	Specifications	Analysis
H ₂ O ₂	20.0 – 24.0%	23.4%
PAA	5.0% ± 1.0%	5.3%
Acetic Acid	8.0 – 10.0%	9.0%
pH, (as is)	0.5 - 1.1 (as is)	0.7

This lot has been reviewed and meets specified requirements.



Quality Analyst



Analysis Date



Quality Assurance AP



Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

RAPICIDE™ is a trademark of Medivators Inc.