



Certificate of Analysis

RAPICIDE™ PA

RTU Part B

Lot Number: 015519
Expiration Date: 2020/10


Components	Specifications	Analysis
Na+	35.5 – 43.4 mEq/L	<u>38.9mEq/L</u>
Benzotriazole	0.05% \pm 0.005%	<u>0.046%</u>
pH, (as is)	10.6 – 11.6 (as is)	<u>11.6</u>

This lot has been reviewed and meets specified requirements.



Quality Analyst

16-APR-19
Analysis Date



Quality Assurance QP

18/04/19
Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

RAPICIDE™ is a trademark of Medivators Inc.