



Certificate of Analysis

RAPICIDE™ PA

RTU Part A

Lot Number: 015619
Expiration Date: 2020/10

Components	Specifications	Analysis
H ₂ O ₂	20.0 – 24.0%	23.4%
PAA	5.0% ± 1.0%	5.3%
Acetic Acid	8.0 – 10.0%	9.0%
pH, (as is)	0.5 - 1.1 (as is)	0.7

This lot has been reviewed and meets specified requirements.

Quality Analyst

Analysis Date

Quality Assurance

Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

RAPICIDE™ is a trademark of Medivators Inc.



Certificate of Analysis

RAPICIDE™ PA

RTU Part B

Lot Number: 015719
Expiration Date: 2020/10

Components	Specifications	Analysis
Na ⁺	35.5 – 43.4 mEq/L	37.3mEq/L
Benzotriazole	0.05% ± 0.005%	0.049%
pH, (as is)	10.6 – 11.6 (as is)	11.6

This lot has been reviewed and meets specified requirements.

FDC

Quality Analyst

17-APR-19

Analysis Date

CO

Quality Assurance

19-APR-19

Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

RAPICIDE™ is a trademark of Medivators Inc.