



Certificate of Analysis

RAPICIDE® PA

Part B

Product Code: ML02-0115
Lot Number: 016118
Expiration Date: 2019/11

Components	Specification	Analysis
Sodium	743.0 - 909.0 mEq/L	882.0 mEq/L
Benzotriazole	0.9 – 1.1% w/w	1.0 %w/w
pH	11.4 to 12.4 as is	11.5

This lot has been reviewed and meets specified requirements.

Quality Analyst

03 May 2018

Analysis Date

04-MAY-18

Release Date

Medivators Inc.
14605 28th Avenue North
Plymouth, Minnesota 55447
Telephone: (763) 553-3300
Fax (763) 553-3387

RAPICIDE® is a registered trademark of Medivators Inc.