



## Certificate of Analysis


# RAPICIDE® PA


### Part B


Product Code: ML02-0115  
Lot Number: 016119  
Expiration Date: 2020/10


Components	Specification	Analysis
Sodium	743.0 - 909.0 mEq/L	778.0 mEq/L
Benzotriazole	0.9 – 1.1% w/w	0.9 %w/w
pH	11.4 to 12.4 as is	11.6

This lot has been reviewed and meets specified requirements.

  
Quality Analyst

  
Analysis Date

  
Quality Assurance

  
Release Date

Medivators Inc.  
14605 28th Avenue North  
Plymouth, Minnesota 55447  
Telephone: (763) 553-3300  
Fax (763) 553-3387

RAPICIDE® is a registered trademark of Medivators Inc.