



Certificate of Conformance

RAPICIDE® OPA-28 Test Strips

Part Number:	73-OPA28B
Vial Lot Number:	017124
Expiration Date:	06/16

Quality Control Report:

Minntech certifies that the test strips indicate a positive or pass reaction with a solution mixed to a concentration of 0.50% \pm 0.05% OPA.

Minntech certifies that the test strips indicate a negative or fail reaction with a solution mixed to a concentration of 0.35% \pm 0.05% OPA.

The statistical analysis and approval test procedure is on file and available for review.

This lot has been reviewed and meets specified requirements.



Quality Assurance

1/5/2015
Release Date

MEDIVATORS Inc.
14605 28th Avenue N
Minneapolis, MN 55447
Tel: 800.444.4729
Fax: 763.553.3387
www.medivators.com