



Certificate of Analysis

RAPICIDE® PA

Part A

Product Code: ML 02-0115
Lot Number: 017517
Expiration Date: 2018/12

Components	Specifications	Analysis
H ₂ O ₂	20.0 – 24.0% (w/w)	23.3 %
PAA	4.0 – 6.0 % (w/w)	5.5 %
Acetic Acid	8.0 – 10.0% (w/w)	8.8 %
pH, (as is)	0.5 – 1.1 (as is)	0.5

This lot has been reviewed and meets specified requirements.

Quality Analyst

06 June 2017

Analysis Date

Quality Assurance

08 June 2017

Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

RAPICIDE® is a registered trademark of Medivators Inc.