



Certificate of Conformance

RAPICIDE® OPA-28 Test Strips

Part Number: **ML02-0137**

Vial Lot Number: **018015A**

Expiration Date: **07/16**

Quality Control Report:

Minntech certifies that the test strips indicate a positive or pass reaction with a solution mixed to a concentration of $0.50\% \pm 0.05\%$ OPA.

Minntech certifies that the test strips indicate a negative or fail reaction with a solution mixed to a concentration of $0.35\% \pm 0.05\%$ OPA.

The statistical analysis and approval test procedure is on file and available for review.

This lot has been reviewed and meets specified requirements.

Quality Assurance

MEDIVATORS Inc.

14605 28th Avenue N

Minneapolis, MN 55447

Tel: 800.444.4729

Fax: 763.553.3387

www.medivators.com

2/18/2015

Release Date