



Certificate of Analysis

RAPICIDE® PA

Part B

Product Code: ML02-0115
Lot Number: 018218
Expiration Date: 2019/11

Components	Specification	Analysis
Sodium	743.0 - 909.0 mEq/L	797.0 mEq/L
Benzotriazole	0.9 – 1.1% w/w	1.0 %w/w
pH	11.4 to 12.4 as is	11.8

This lot has been reviewed and meets specified requirements.

PS

Quality Analyst

21 May 2018

Analysis Date

Quality Assurance

23-MAY-18

Release Date

Medivators Inc.

14605 28th Avenue North

Plymouth, Minnesota 55447

Telephone: (763) 553-3300

Fax (763) 553-3387

RAPICIDE® is a registered trademark of Medivators Inc.