



Certificate of Analysis


RAPICIDE® PA

RTU Part B

Lot Number: 018517
Expiration Date: 2018/12
Quality Control Report: 170596

Components	Specifications	Analysis
Na+	35.5 – 43.4 mEq/L	40.1
Benzotriazole	0.05% \pm 0.005%	0.049
pH, (as is)	10.6 – 11.6 (as is)	11.6


This lot has been reviewed and meets specified requirements.



Quality Analyst 

14 June 2017
Analysis Date



Quality Assurance 

19 June 2017
Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com