



Certificate of Analysis

RAPICIDE™ PA

RTU Part A

Lot Number: 018518
Expiration Date: 2019/11

Components	Specifications	Analysis
H ₂ O ₂	20.0 – 24.0%	<u>22.8%</u>
PAA	5.0% ± 1.0%	<u>5.1%</u>
Acetic Acid	8.0 – 10.0%	<u>9.2%</u>
pH, (as is)	0.5 - 1.1 (as is)	<u>0.9</u>

This lot has been reviewed and meets specified requirements.

PSC
Quality Analyst 23 May 18
Analysis Date

QA
Quality Assurance 23 May 18
Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

RAPICIDE™ is a trademark of Medivators Inc.