

Certificate of Analysis

RAPICIDE® PA

Part B

Product Code: **ML02-0115**
Lot Number: **018716**
Expiration Date: **02/2018**

Components	Test Method	Specification	Analysis
Sodium	20100-070	743.0 - 909.0 mEq/L	827.5 mEq/L
Benzotriazole	20601-489	0.9 – 1.1% w/w	1.0 %w/w
pH	20002-154	11.4 to 12.4 as is	11.5

This lot has been reviewed and meets specified requirements.



Quality Analyst

30Aug2016

Analysis Date



Quality Assurance

30Aug2016

Release Date

Medivators Inc.
14605 28th Avenue North
Plymouth, Minnesota 55447
Telephone: (763) 553-3300
Fax (763) 553-3387

RAPICIDE® is a registered trademark of Medivators Inc.