



Quality Control Report and Certificate of Analysis

Formula: MB-330-25

Lot #: 01NN08

Concentrate Specification

Sodium 967.2 mEq/L $\pm$ 2 %

Bicarbonate 967.2 mEq/L $\pm$ 2 %

pH 7.0 - 8.0 pH units

Conductivity @ 25 °C 44.5 $\pm$ 2 mS/cm

When mixed using the correct equipment and dilution with AAMI certified water MB-330-25 ,
Lot # 01NN08 , meets the above specifications.

Approved for Release

Quality Control

Elaine Trogen

Date:

02/06/08