



Quality Control Report and Certificate of Analysis

Formula: MB-330-25

Lot #: 01VV08

Concentrate Specification

Sodium	<u>967.2</u> mEq/L \pm 2 %
Bicarbonate	<u>967.2</u> mEq/L \pm 2 %
pH	<u>7.0 - 8.0</u> pH units
Conductivity @ 25 °C	<u>44.5</u> \pm 2 mS/cm

When mixed using the correct equipment and dilution with AAMI certified water MB-330-25 ,
Lot # 01VV08 , meets the above specifications.

Approved for Release

Quality Control APL

Date: 9-5-08