



Quality Control Report and Certificate of Analysis

Formula: BC-1-25
Lot #: 01WW07

Concentrate Specification

Sodium	<u>1187.7</u> mEq/L \pm 2 %
Chloride	<u>402.6</u> mEq/L \pm 2 %
Bicarbonate	<u>785.1</u> mEq/L \pm 2 %
pH	<u>7.0 - 8.0</u> pH units
Conductivity @ 25 °C	<u>67</u> \pm 3 mS/cm

When mixed using the correct equipment and dilution with AAMI certified water BC-1-25,
Lot # 01WW07, meets the above specifications.

Approved for Release

Quality Control Elaine Troje Date: 09/06/07
20600-379/D