



Quality Control Report and Certificate of Analysis

Formula: MB-330-15

Lot #: 01YY07

Concentrate Specification

Sodium	<u>967.2</u> mEq/L $\pm$ 2 %
Bicarbonate	<u>967.2</u> mEq/L $\pm$ 2 %
pH	<u>7.0 - 8.0</u> pH units
Conductivity @ 25 °C	<u>44.5</u> $\pm$ 2 mS/cm

When mixed using the correct equipment and dilution with AAMI certified water MB-330-15 ,
Lot # 01YY07 , meets the above specifications.

Approved for Release

Quality Control

A handwritten signature in black ink, appearing to read "APB", followed by a horizontal line.

Date:

11-6-07