



## DOCUMENTATION

DOCUMENT  
NUMBER

20100-706

REVISION: E  
PAGE 7 OF 7

## Quality Control report and certificate of analysis

Formula: SB-1003-R  
Lot #: 01207

<u>Component</u>	<u>Concentrate Specification</u>
Sodium	<u>2946.4</u> mEq/l $\pm$ 2%
Calcium	<u>128.9</u> mEq/l $\pm$ 4%
Potassium	<u>73.7</u> mEq/l $\pm$ 2%
Magnesium	<u>36.8</u> mEq/l $\pm$ 4%
Chloride	<u>3185.8</u> mEq/l $\pm$ 2%
Dextrose	<u>7366</u> mg/dl $\pm$ 4%
pH (Conc.)	<u>2.0 – 3.0</u>
pH (Dial.)	<u>6.5 – 7.4</u>
Conductivity 25°C (Conc.)	<u>151.5</u> mS/cm $\pm$ 5%
Conductivity 25°C (Dial.)	<u>13.95</u> mS/cm $\pm$ 5%

When mixed using the correct equipment and dilution with AAMI certified water Formula SB-1003-R,  
Lot # 01207, meets the above specifications.

Approved for release:

  
\_\_\_\_\_  
Quality Control

12-14-07

\_\_\_\_\_  
Date

RELEASED  
18-May-2005

THIS DOCUMENT AND ALL INFORMATION CONTAINED HEREIN IS THE PROPERTY OF MINNTECH CORPORATION AND IS PROPRIETARY AND CONFIDENTIAL. THIS CONFIDENTIAL INFORMATION MUST NOT BE MADE PUBLIC, DISCLOSED TO ANY THIRD PARTY OR COPIED UNLESS AUTHORIZED BY AN OFFICER OF MINNTECH, AND IS SUBJECT TO RETURN UPON DEMAND