

Certificate of Conformance

RAPICIDE® OPA-28 Test Strips

Part Number:

ML02-0136

Vial Lot Number:

020045

Expiration Date:

12/16

Quality Control Report:

Minntech certifies that the test strips indicate a positive or pass reaction with a solution mixed to a concentration of $0.50\% \pm 0.05\%$ OPA.

Minntech certifies that the test strips indicate a negative or fail reaction with a solution mixed to a concentration of $0.35\% \pm 0.05\%$ OPA.

The statistical analysis and approval test procedure is on file and available for review.

This lot has been reviewed and meets specified requirements.



Quality Assurance

4/24/2015

Release Date

MEDIVATORS Inc.

14605 28th Avenue N

Minneapolis, MN 55447

Tel: 800.444.4729

Fax: 763.553.3387

www.medivators.com