



**MINNTECH®**  
A Cantel Medical Company

*Certificate of Conformance*  
**RAPICIDE TEST STRIP**

**Part Number:** ML02-0120

**Lot Number:** 020332

**Expiration Date:** 2012-12

I certify that the Rapicide test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the Rapicide test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.



Quality Analyst

03/02/11

Analysis Date



Quality Assurance

03/02/11

Release Date

**MINNTECH CORPORATION**  
14605 28th Avenue North  
Plymouth, Minnesota 55447  
Telephone: (763) 553-3300  
Fax (763) 553-3387