

# Certificate of Analysis

## **RAPICIDE® PA**

### **Part B**

Product Code: **ML02-0115**

Lot Number: **020818**

Expiration Date: **2019/12**

Components	Specification	Analysis
Sodium	743.0 - 909.0 mEq/L	787.0 mEq/L
Benzotriazole	0.9 – 1.1% w/w	0.9 %w/w
pH	11.4 to 12.4 as is	11.5

This lot has been reviewed and meets specified requirements.



Quality Analyst

20 June 2018

Analysis Date



Quality Assurance

20 Jun 18

Release Date

**Medivators Inc.**

14605 28th Avenue North  
Plymouth, Minnesota 55447

Telephone: (763) 553-3300

Fax (763) 553-3387

RAPICIDE® is a registered trademark of Medivators Inc.