



MINNTECH®
A Cantel Medical Company

Certificate of Conformance

RAPICIDE TEST STRIP

Part Number: ML02-0120

Lot Number: 020886

Expiration Date: 2013/04

I certify that the Rapicide test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the Rapicide test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.

Jennifer Osterhamer
Quality Analyst

5/11/11
Analysis Date

Stephen Orvis
Quality Assurance

5/11/11
Release Date

MINNTECH CORPORATION
14605 28th Avenue North
Plymouth, Minnesota 55447
Telephone: (763) 553-3300
Fax (763) 553-3387