



# **MINNTECH®**

---

**A Cantel Medical Company**

## ***Certificate of Conformance***

# **RAPICIDE TEST STRIP**

**Part Number:** ML02-0120

**Lot Number:** 020988

**Expiration Date:** 2013/04

I certify that the Rapicide test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the Rapicide test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.

  
Jennifer Ostreich Farmer  
Quality Analyst

5/23/11  
Analysis Date

Stahn Reich  
Quality Assurance

5/23/11

**MINNTECH CORPORATION**  
14605 28th Avenue North  
Plymouth, Minnesota 55447  
Telephone: (763) 553-3300  
Fax (763) 553-3387