



# MEDIVATORS

A Cantel Medical Company

# Certificate of Analysis

# **RAPICIDE® PA**

## RTU Part B

Lot Number: 021016  
Expiration Date: 2018/03  
Quality Control Report:

Components	Specifications	Analysis
Na <sup>+</sup>	35.5 – 43.4 mEq/L	42.0 mEq/L
Benzotriazole	0.05% $\pm$ 0.005%	0.049%
pH, (as is)	10.6 – 11.6 (as is)	11.4

This lot has been reviewed and meets specified requirements.

Andee Hudecker 27Sep2016  
Quality Analyst Analysis Date

*Erin D. Gootz* 27Sep2016  
Quality Assurance Release Date

**Medivators Inc.**  
14605 28th Avenue North  
Minneapolis, MN 55447 USA  
Telephone: (763) 553-3300  
Fax (763) 553-3387  
[www.medivators.com](http://www.medivators.com)