



Certificate of Conformance

RAPICIDE TEST STRIP

Part Number: ML02-0120


Lot Number: 021211

Expiration Date: 2013/04

I certify that the Rapicide test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the Rapicide test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.



Quality Analyst

06/23/11
Analysis Date



Jennifer Ostreich-Farmer
Quality Assurance

06/23/11
Release Date

MINNTECH CORPORATION
14605 28th Avenue North
Plymouth, Minnesota 55447
Telephone: (763) 553-3300
Fax (763) 553-3387