



MINNTECH®
A Cantel Medical Company

Certificate of Conformance

RAPICIDE TEST STRIP

Part Number: ML02-0120

Lot Number: 021471

Expiration Date: 2013-09

I certify that the Rapicide test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the Rapicide test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.

Quality Analyst

09/23/11

Analysis Date

Quality Assurance

09/23/11

Release Date

MINNTECH CORPORATION

14605 28th Avenue North
Plymouth, Minnesota 55447
Telephone: (763) 553-3300
Fax (763) 553-3387