



Certificate of Conformance

RAPICIDE TEST STRIP

Part Number: ML02-0120

Lot Number: 022001

Expiration Date: 2013-12

I certify that the Rapicide test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the Rapicide test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.

02/02/12

Analysis Date

Quality Analyst

Quality Assurance

02/02/12

Release Date

MINNTECH CORPORATION

14605 28th Avenue North

Plymouth, Minnesota 55447

Telephone: (763) 553-3300

Fax (763) 553-3387