



## *Certificate of Conformance*

# **RAPICIDE TEST STRIP**

**Part Number:** ML02-0120

**Lot Number:** 022234

**Expiration Date:** 2014-01

I certify that the Rapicide test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the Rapicide test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.

Quality Analyst

02/28/12

Analysis Date

Quality Assurance

02/28/12

Release Date

### **MINNTECH CORPORATION**

14605 28th Avenue North  
Plymouth, Minnesota 55447

Telephone: (763) 553-3300  
Fax (763) 553-3387