



Certificate of Analysis

RAPICIDE® PA

Part B

Product Code: ML02-0115
Lot Number: 022317
Expiration Date: 2019/01


Components	Specification	Analysis
Sodium	743.0 - 909.0 mEq/L	804.9 mEq/L
Benzotriazole	0.9 – 1.1% w/w	1.1 %w/w
pH	11.4 to 12.4 as is	11.7

This lot has been reviewed and meets specified requirements.



Quality Analyst

24 July 2017
Analysis Date



Quality Assurance

26 July 2017
Release Date

Medivators Inc.
14605 28th Avenue North
Plymouth, Minnesota 55447
Telephone: (763) 553-3300
Fax (763) 553-3387

RAPICIDE® is a registered trademark of Medivators Inc.