



## *Certificate of Conformance*

# **RAPICIDE TEST STRIP**

**Part Number:** ML02-0120

**Lot Number:** 022821

**Expiration Date:** 2014/07

I certify that the Rapicide test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the Rapicide test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.


The statistical analysis and approved test procedure is on file and available for review.

A handwritten signature in blue ink, appearing to read "Leah".

Quality Analyst

09/17/12

Analysis Date

A handwritten signature in blue ink, appearing to read "Elaine Troge".

Quality Assurance

09/17/12

Release Date

**MINNTECH CORPORATION**

14605 28th Avenue North

Plymouth, Minnesota 55447

Telephone: (763) 553-3300

Fax (763) 553-3387