



Certificate of Conformance

RAPICIDE TEST STRIP

Part Number:	ML02-0120
Lot Number:	023125
Expiration Date:	2014-08

I certify that the Rapicide test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the Rapicide test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.

A handwritten signature in black ink.

Quality Analyst

10/16/12

Analysis Date

A handwritten signature in black ink.

Quality Assurance

10/16/12

Release Date

MINNTECH CORPORATION

14605 28th Avenue North
Plymouth, Minnesota 55447
Telephone: (763) 553-3300
Fax (763) 553-3387