



MINNTECH®
A Cantel Medical Company

Certificate of Conformance

RAPICIDE TEST STRIP

Part Number: ML02-0120

Lot Number: 023462

Expiration Date: 2014/10

I certify that the Rapicide test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the Rapicide test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.



Quality Analyst

11/21/12

Analysis Date



Quality Assurance

11/21/12

Release Date

MINNTECH CORPORATION
14605 28th Avenue North

Plymouth, Minnesota 55447

Telephone: (763) 553-3300

Fax (763) 553-3387