



## Certificate of Analysis

# RAPICIDE® PA

### Part A

Product Code:

**ML02-115**

Lot Number:

023818

Expiration Date:

2020/01

| Components                    | Specifications     | Analysis |
|-------------------------------|--------------------|----------|
| H <sub>2</sub> O <sub>2</sub> | 20.0 – 24.0% (w/w) | 22.5 %   |
| PAA                           | 4.0 – 6.0 % (w/w)  | 5.0 %    |
| Acetic Acid                   | 8.0 – 10.0% (w/w)  | 9.3 %    |
| pH, (as is)                   | 0.5 – 1.1 (as is)  | 0.6      |

This lot has been reviewed and meets specified requirements.

Quality Analyst

18 July 2018

Analysis Date

Quality Assurance

201 July 2018

Release Date

Medivators Inc.  
14605 28th Avenue North  
Minneapolis, MN 55447 USA  
Telephone: (763) 553-3300  
Fax (763) 553-3387  
[www.medivators.com](http://www.medivators.com)

RAPICIDE® is a registered trademark of Medivators Inc.