

Certificate of Conformance

RAPICIDE® OPA/28 Test Strips

Part Number: **ML02-0136**

Lot Number: 024026

Vial Lot Number: 031026

Expiration Date: 10/17

Quality Control Report:

MEDIVATORS certifies that the test strips indicate a positive or pass reaction with a solution mixed to a concentration of $0.50\% \pm 0.05\%$ OPA.

MEDIVATORS certifies that the test strips indicate a negative or fail reaction with a solution mixed to a concentration of $0.35\% \pm 0.05\%$ OPA.

The statistical analysis and approval test procedure is on file and available for review.

This lot has been reviewed and meets specified requirements.


Quality Assurance

4/20/2016
Release Date

MEDIVATORS Inc.
14605 28th Avenue N
Minneapolis, MN 55447
Tel: 800.444.4729
Fax: 763.553.3387
www.medivators.com