



## Certificate of Analysis

# RAPICIDE® PA

### Part B

Product Code: ML02-0115  
Lot Number: 024118  
Expiration Date: 2020/01

Components	Specification	Analysis
Sodium	743.0 - 909.0 mEq/L	828.0 mEq/L
Benzotriazole	0.9 – 1.1% w/w	1.1 %w/w
pH	11.4 to 12.4 as is	11.6

This lot has been reviewed and meets specified requirements.

*PS*  
\_\_\_\_\_  
Quality Analyst  
  
*QC*  
\_\_\_\_\_  
Quality Assurance

19 July 2018  
\_\_\_\_\_  
Analysis Date

*13-JUL-18*  
\_\_\_\_\_  
Release Date

Medivators Inc.  
14605 28th Avenue North  
Plymouth, Minnesota 55447  
Telephone: (763) 553-3300  
Fax (763) 553-3387

RAPICIDE® is a registered trademark of Medivators Inc.