



## *Certificate of Conformance*

# **RAPICIDE TEST STRIP**

**Part Number:** ML02-0120

**Lot Number:** 024347

**Expiration Date:** 2015-05

I certify that the Rapicide test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the Rapicide test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.

Quality Analyst

7/24/2013

Analysis Date

Quality Assurance

7/24/2013

Release Date

**MINNTECH CORPORATION**  
14605 28th Avenue North  
Plymouth, Minnesota 55447  
Telephone: (763) 553-3300  
Fax (763) 553-3387