



Certificate of Analysis

RAPICIDE® PA

Part A

Product Code: ML02-0115
Lot Number: 024416
Expiration Date: 2018/04

Components	Specifications	Analysis
H ₂ O ₂	20.0 – 24.0% (w/w)	23.3 %
PAA	4.0 – 6.0 % (w/w)	5.3 %
Acetic Acid	8.0 – 10.0% (w/w)	9.3 %
pH, (as is)	0.5 – 1.1 (as is)	0.6

This lot has been reviewed and meets specified requirements.

Quality Analyst

25 Oct 2016

Analysis Date

Quality Assurance

26 Oct 2016

Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

RAPICIDE® is a registered trademark of Medivators Inc.