



## *Certificate of Conformance*

# RAPICIDE TEST STRIP

**Part Number:** ML02-0120


**Lot Number:** 024466

**Expiration Date:** 2015-06

I certify that the Rapicide test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the Rapicide test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.

  
Quality Analyst

26AUG2013

Analysis Date

  
Quality Assurance

26AUG2013

Release Date

**MINNTECH CORPORATION**  
14605 28th Avenue North  
Plymouth, Minnesota 55447  
Telephone: (763) 553-3300  
Fax (763) 553-3387