



## Certificate of Analysis

# RAPICIDE™ PA

### RTU Part A

Lot Number: 024518  
Expiration Date: 2020/01

Components	Specifications	Analysis
H <sub>2</sub> O <sub>2</sub>	20.0 – 24.0%	22.4%
PAA	5.0% ± 1.0%	5.5%
Acetic Acid	8.0 – 10.0%	8.8%
pH, (as is)	0.5 - 1.1 (as is)	0.8

This lot has been reviewed and meets specified requirements.

PS

Quality Analyst

25 July' 18

Analysis Date

  
Quality Assurance

27 JUL 18  
Release Date

Medivators Inc.  
14605 28th Avenue North  
Minneapolis, MN 55447 USA  
Telephone: (763) 553-3300  
Fax (763) 553-3387  
[www.medivators.com](http://www.medivators.com)

RAPICIDE™ is a trademark of Medivators Inc.