



Certificate of Analysis

RAPICIDE™ PA

RTU Part A

Lot Number: 025018
Expiration Date: 2020/01


Components	Specifications	Analysis
H ₂ O ₂	20.0 – 24.0%	<u>22.5%</u>
PAA	5.0% ± 1.0%	<u>5.3%</u>
Acetic Acid	8.0 – 10.0%	<u>9.0%</u>
pH, (as is)	0.5 - 1.1 (as is)	<u>0.8</u>

This lot has been reviewed and meets specified requirements.



Quality Analyst 26 July' 18

Analysis Date



Quality Assurance 30 JUL 18

Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

RAPICIDE™ is a trademark of Medivators Inc.