



MINNTECH®
A Cantel Medical Company

Certificate of Conformance

RAPICIDE TEST STRIP

Part Number: ML02-0120

Lot Number: 025033

Expiration Date: 2016-01

I certify that the Rapicide test strips indicate a positive reaction with a solution mixed to a concentration of 2.5% glutaraldehyde.

I certify that the Rapicide test strips indicate a negative reaction with a solution mixed to a concentration of 1.3% glutaraldehyde.

The statistical analysis and approved test procedure is on file and available for review.

Quality Analyst

03FEB2014

Analysis Date

Quality Assurance

03FEB2014

Release Date

MINNTECH CORPORATION

14605 28th Avenue North
Plymouth, Minnesota 55447
Telephone: (763) 553-3300
Fax (763) 553-3387